

Prevent Electric Shock Probe Power Indicator



Input voltage: AC 220V, 50Hz, 1A

Output voltage: AC 42V, 50Hz, 25mA

The standard test finger is applied in every possible position, an electrical indicator with a voltage between 40V and 50V being used to show contact with the relevant parts.

